

Consultative Meeting Notification 124/ 2017

**Buyer-seller consultation on Biopsy Needle dated
19-12-2017**

Presented by :

Date :



Consultative Meeting Notification 124/ 2017

Buyer-seller consultation on Biopsy Needle dated 19-12-2017

Consultative Meeting held on 19th December for BIOPSY NEEDLE with sellers and buyers. Following participated in the meeting:

Buyers participants:

1. AIIMS, New Delhi
2. RML Hospital
3. LHMC & SSKH
4. Defence, Assam Rifles
5. GNCTD Bhagwan Mahavir Hospital
6. Deepchandra Hospital

Sellers participants:

1. Medax/Hospitek marketing
 2. Kshemin Consultant
 3. Salveo Medical Systems
 4. Axio Biosolution
 5. Axio Biosolution pvt. Ltd. Ahmedabad, OEM
 6. Spanco Industries
 7. Vaishnvi Med. Eq. System
 8. BVM Medical System
 9. Swastik Traders
-
1. Following **Golden parameters** were decided based on the suggestions and discussions –
 - Type of bone marrow biopsy needle
 - Patient category
 - Needle size
 - Double diamond stylet point
 - Twist lock handle
 2. Following changes were decided based on the suggestions and discussions-
 - A type of bone marrow biopsy needle should be with/without T-handle needle.
 - Adjustable length to be deleted



- Compatible with luer slip and luer lock must be mandatory
- Aspiration needle must be included
- Method of sterilization may be ETO/Gamma
- Manufacturers Drug License number and date to be included
- Shelf life of 5 years to be included
- Cap of needle may be made optional by the supplier
- Length of needle may be 50,70, 100, 150mm included
- A needle size of 18G may be added.

Regards

GeM-Admin