

Consultative Meeting Notification 22/2017

Buyer-seller consultation on Central Venous Catheters of single,
Double, Triple Lumen dated 24-11- 2017

Presented by :

Date :





Minutes of the Consultative meeting held on 24.11.2017 at 11.30 PM under Chairmanship of Shri A V Muralidharan Dy CEO (GeM) at the Office of Government e Market (SPV) , New Delhi regarding Medical items

The following attended Meeting from seller side:

1. Ms Life Long Meditech pvt. Ltd.
2. M/s Hindustan Syringe Ltd
- 3 . M/s B. Braun Medical (I) Pvt Ltd
4. M/s B.Brawn Medical (I) Pvt Ltd
5. M/s Mediplus India Ltd
6. M/s Polymedicure Medical Devices
7. M/s Sterimed Medical Devices Pvt Ltd
8. M/s ISCON Surgicals ltd
9. M/s Syna india associates
10. M/s Alliance Medi Equipment Ltd

Buyer side:

None Present.

Based on suggestion received in the meeting following were decided to be included in existing TP

- ☐ Option for quad and pent Lumen in existing TP
- ☐ More sizes i.e. 3, 4, 4.5, 5.5 in existing TP
- ☐ Shape of introducer shape of needle i.e straight or y shaped
- ☐ 5ml in capacity of syringes
- ☐ Guidewire dispenser material ABS type with yes / no option
- ☐ Solo connect availability option with yes /no in existing TP
- ☐ Drug license requirements
- ☐ Type of coating



- ☐ Shelf life and confirmation of having at least 2/3 of shelf life available at time of supplies
- ☐ Certifications such as Good Manufacturing Practices (GMP) under revised Schedule -M or WHO-Good Manufacturing Practices (GMP) or certifications such as CE/FDA/CSA/PQS in respect of the original manufacturer
- ☐ Requirement of Non conviction certificate.
- ☐ Requirement of batch wise in house test reports copies.

Golden Parameter proposed by participants are as under

- ☐ Radio Opaque nature
- ☐ Size
- ☐ Length of Catheters
- ☐ Type of lumen
- ☐ Type of coating

Deletions from existing TP

No deletions from existing TP and only additions required

Regards

Gem-Admin

