

## Consultative Meeting Notification 125/2017

Buyer-seller consultation on Continuous Plexus Anaethesia Sets Anesthesia Nerve Block Trays or Needles dated 19-12-2017

Presented by :

Date :



Consultative Meeting Notification 125/2017

## Buyer-seller consultation on CONTINUOUS PLEXUS ANAETHESIA SETS ANESTHESIA NERVE BLOCK TRAYS OR NEEDLES dated 19-12-2017

Consultative Meeting held on 19<sup>th</sup> December for CONTINUOUS PLEXUS ANAETHESIA SETS ANESTHESIA NERVE BLOCK TRAYS OR NEEDLES with sellers and buyers. Following participated in the meeting:

## Buyers participants:

- 1. AIIMS, New Delhi
- 2. RML Hospital
- 3. LHMC & SSKH
- 4. Defence, Assam Rifles
- 5. GNCTD Bhagwan Mahavir Hospital
- 6. Deepchandra Hospital

## Sellers participants:

- 1. Medax/Hospitek marketing
- 2. Kshemin Consultant
- 3. Salveo Medical Systems
- 4. Axio Biosolution
- 5. Axio Biosolution pvt. Ltd. Ahmedabad, OEM
- 6. Spanco Industries
- 7. Vaishnvi Med. Eq. System
- 8. BVM Medical System
- 9. Swastik Traders

1. Following **Golden parameters** were decided based on the suggestions and discussions –

- Size (length of needle)
- 2. Following changes were decided based on the suggestions and discussions:
  - Pre-sterilization and method of sterilization to be made mandatory



- Needle should have reflector or visualization during ultra-sound—to be made mandatory
- Manufacturers Drug License number and date to be included
- Shelf life of 5 years to be included
- Availability of test report from Central Govt./ NABL/ILAC may be replaced with submission of test report on Form No.39 or from Central Govt./ NABL/ILAC along with supply to the buyer
- All certification to be provided to the buyer on demand

Regards GeM-Admin