

Consultative Meeting Notification 21/2017

Buyer-seller consultation on IV Cannula dated 24-11- 2017

Presented by :

Date :





Minutes of the Consultative meeting held on 24.11.2017 at 11.30 PM under Chairmanship of Shri A V Muralidharan Dy CEO (GeM) at the Office of Government e Market (SPV) New Delhi regarding Medical items

The following attended Meeting from seller side:

1. Ms Life Long Meditech Pvt. Ltd.
2. M/s Hindustan Syringe Ltd
- 3 . M/s B. Braun Medical (I) Pvt Ltd
4. M/s B.Brawn Medical (I) Pvt Ltd
5. M/s Mediplus India Ltd
6. M/s Polymedicure Medical Devices
7. M/s Sterimed Medical Devices Pvt Ltd
8. M/s ISCON Surgicals Ltd
9. M/s Synaindia associates Ltd
10. M/s Alliance Medi Equipment Ltd

Buyer side:

None Present.

Based on suggestions received in the meeting following were decided to be included in existing TP

- ☐ IV cannula specification as IS/ISO 10555-5
- ☐ Parameter nominal outside diameter of catheter tube and color coding as per IS/ISO 10555-5
- ☐ Needle point finishing as long bevel cut or short bevel cut
- ☐ Parameter fitting with 6% luer Taper with yes or no option .
- ☐ Needle hub as complying with ISO 594-1
- ☐ Radio opaque line features with yes / no option.



- ☐ I V cannulas with safety features for preventing needle stick injuries
- ☐ Drug License provisions
- ☐ Shelf life and confirmation of having at least 2/3 of shelf life available at time of supplies
- ☐ Certifications such as Good Manufacturing Practices (GMP) under revised Schedule -M or WHO-Good Manufacturing Practices (GMP) or certifications such as CE/FDA/CSA/PQS in respect of the original manufacturer
- ☐ Requirement of Non conviction certificate.
- ☐ Requirement of batch wise in house test reports copies.

Golden Parameters Proposed by Participants are as under

- ☐ Gauge of the IV Canulla
- ☐ ,Material of IV Canulla
- ☐ Whether radio opaque or not
- ☐ Whether safety features for needle stick injuries is available.
- ☐ Bevel of the needle

After deliberations with the participants, it was decided the following parameters are to be deleted from the existing TP:

- ☐ Option of without catheter
- ☐ outer diameter of needle and nominal outside diameter of catheter
- ☐ The requirement of Test Reports.

Regards

GeM-Admin



GeM
Government
e Marketplace