

Consultative Meeting Notification 31 /2017

**Buyer-Seller consultation on Instrument Trolleys for
Hospitals dated 23.11.2017**

Presented by :

Date :



Consultative Meeting Notification 31 /2017

Buyer-Seller consultation on Instrument Trolleys for hospitals

dated 23.11.2017

Meeting held on 23th November 2017 for Instrument Trolleys with sellers and Buyers.

Following participated from the seller side:

M/s Vinar System Private Limited.

M/s S.S Enterprises

M/s Perfect Supplier

M/s Continental Hygiene Concepts

Buyer side:

None Present. However input received from Dr. RML Hospital on Email.

In the consultative meeting the following points are decided by participants.

- Instrument trolleys for hospitals renamed to trolleys for hospitals items like oxygen trolleys, medicine trolley, dressing trolleys and instrument trolleys as input received from RML Hospitals
- Overall size of trolley kept open for declaration by vendors
- Add foldable or not foldable option
- Materials to be specified such as MS, SS201, SS202, SS304, carbon steel, fabric, plastic and others
- To include load carrying capacity parameter must be there in TP.
- Thickness of sheet must be declared by vendors itself varying from 0.5 mm to 5 mm intervals of 0.5 mm
- Declaration of ground clearance must be declared by vendors itself
- Conformity to Indian standards for trolley if any
- Nature of operation as manual or motorized must be mentioned



- volume carrying capacity of trolley
- Corrosive or non corrosive environmental use of operation
- Number of shelves as 0, 1, 2, 3, 4
- Multiple shelves thickness of sheet must be uniform.
- Wheel size to be declared , for number of wheel option 2, 3, 4
- Type of wheel movement as in existing TP
- In addition of powder coating NA option must be there for materials other than MS.
- Railing option in TP as one side, two side, three side, four side

Golden Parameter for Instrument Trolleys for hospitals:

- Material of Trolleys
- Load Carrying capacity
- Overall Dimensions of trolleys
- Type of environment in which Trolley used
- Thickness of Sheet

Deletion to be made for Instrument Trolleys for hospitals:

- Nil

Regards

GeM-Admin