

Consultative Meeting Notification 24/2017

Buyer-seller consultation on Intravenous or Arterial tubing adopters or connectors-plastic connectors (3 way stop cock without tube) dated 24-11-17

Minutes of the Consultative meeting held on 24.11.2017 at 11.30 PM under Chairmanship of Shri A V Muralidharan Dy CEO (GeM) at ISQM Hall ,Office of Government e Market (SPV) New Delhi regarding Medical items;

The following attended Meeting from seller side:

1. Ms Life Long Meditech pvt. Ltd.
2. M/s Hindustan Syringe Ltd
- 3 . M/s B. Braun Medical (I) Pvt Ltd
4. M/s B.Brawn Medical (I) Pvt Ltd
5. M/s Mediplus India Ltd
6. M/s Polymedicure Medical Devices
7. M/s Sterimed Medical Devices Pvt Ltd
8. M/s ISCON Surgicals Ltd
9. M/s Syna india associates
10. M/s Alliance Medi Equipment Ltd

Buyer side:

None Present.

Based on suggestions received in the meeting following were decided to be included in existing TP:

- Decided to rename 3 way stop cock without tubing in view of Difficult to identify the product under this heading
- Arrow option to indicate the flow of liquid
- Options of material as polypropylene and clear polycarbonate
- For silver ion protection provide options yes/ no in exiting TP

- Material of the item
- lipid resistance or not
- Pressure rating in bar /psi
- clear housing / non clear housing option

Golden Parameters Proposed by Participants are as under

- Length of connector
- Material
- Diameter
- Lipid Resistance

After deliberations with the participants, it was decided the following parameters are to be deleted from the existing TP:

- Remove y shape option available in Existing TP in parameter no 1
- Remove various options available for material and keep only 2 option polypropylene and clear polycarbonate.

Regards

Gem-Admin