

## Consultative Meeting Notification 42/2017

**Buyer-seller consultation on Intubation Laryngoscopes –  
Laryngeal Tube with Drainage Tube dated 06.12.2017**

Presented by :

Date :



## Consultative Meeting Notification 42/2017

### **Buyer-seller consultation on INTUBATION LARYNGOSCOPES – LARYNGEAL TUBE WITH DRAINAGE TUBE dated 06.12.2017**

Consultative Meeting held on 06<sup>th</sup> December 2017 with sellers and buyers. Following participated in the meeting:

#### **Sellers Participants**

1. M/s HOSPIMAX
2. M/s AMBU
3. M/s Track Manufacturing
4. M/s La-med healthcare (P) Ltd
5. M/s Hansraj Nayyar Medical
6. M/s Syana Health care
7. M/s Medexcel impex pvt. Ltd.

#### **Buyers Participants**

1. AIIMS
2. SAFDARJUNG HOSPITAL
3. RML HOSPITAL
4. LHMC & SSKH
5. SDDMSC
6. LBS HOSPITALS

1. Following **Golden parameters** were decided based on the suggestions and discussions –
  - a) Material of Tube
  - b) Inner Diameter (I.D)
  - c) Method of sterilization
2. Following changes were decided in the existing TP based on the suggestions and discussions-
  - a) Material of the Tube ( Non-toxic, Medical Grade)– 100% Silicon, 100% PVC and Silicon coated and Others to be added
  - b) Size - “0” is to be added
  - c) Lumen tube - “dual” only retained and option ‘Single’ is to be removed
  - d) Tube to be kink resistance and transparent- Yes to be added
  - e) I.D of the tube clearly marked on the tube – Mandatory
  - f) Utility should be single use only not reusable
  - g) Method of Sterilization- (1) ETO, (2) Gamma and (3) Others to be added.



- h) Tissue implantation Testing certificate – Mandatory
- i) Certification – Must Declare ISO/European CE/US-FDA
- j) Manufacturing Drug License- Yes
- k) Authority of Drug License- Must Declare
- l) Drug license No. - Must Declare
- m) Drug license Date- Must Declare
- n) The product should have shelf life of at least 2/3 rd of the total shelf life at the time of supply
- o) All certification to be provided to the buyer on demand

Regards

GeM-Admin