

Consultative Meeting Notification 84/2017

Buyer-Seller Consultation on Medical/Laboratory Equipment Items dated 13.12.2017

Presented by :

Date :



Consultative Meeting Notification 84/2017

**Buyer-Seller consultation on Medical/Laboratory equipment items dated
13.12.2017**

Consultative Meeting held on 13th December 2017 for Medical/Laboratory equipment items with sellers and buyers by DCEO(GS). Following participants participated from the

Sellers' side:

- (i) CHSPL Common Health Pvt Ltd
- (ii) Karcia Instruments & Calibration Services
- (iii) Grome Medical
- (iv) BMS Health Aids
- (v) Bluemount Appliances
- (vi) Syana India
- (vii) KSI Medicare P Ltd
- (viii) Metro HM India Ltd

Buyers' side:

Nil

In the consultative meeting the following points are suggested by participants:

HOSPITAL BEDS WITH MECHANICAL FITTINGS i.e. Patient Care Bed for General Use:

- (a) Replace phrase "sheet bed" with "mattress platform" and other typographical correction in category name and else where.
- (b) Defining length in terms of two parameters: overall length and mattress platform length, which would be interrelated conditional parameters with max.difference of 100 mm over the mattress platform length within range of 1900 to 2100 mm.
- (c) Height of mattress platform, 650 mm, tolerance $\pm 5\%$
- (d) Height of head/Leg board from platform of mattress : 350 mm -450 mm with Pipe with thickness of 1.2 mm and above, pipe dia. Of 30 mm or more
- (e) Standardisation of various dimensional parameter; length, width(≥ 900 mm), and height of bed, diameter and thickness of pipes used
- (f) Enhancing the Material options for head/leg board as SS (laminated), MS (tubular) and Plastic.



- (g) Material of the frame as MS ERW rectangular tube with 50 mmx30 mm or more, gauge 1.2 mm or more.
- (h) Defining Size of the frame as length and breadth both
- (i) Request for incorporating electro-mechanical/hydraulic beds as separate category to be considered later after obtaining inputs from stakeholders in future.
- (j) Discontinuance of parameter “No. of Holes in Lying Sheet” in TP
- (k) Deletion of “Stove enamelled” option against painting parameter.
- (l) Provision of max. Load bearing capacity as ≥ 150 kg.
- (m) Provision of I.V. poles Holder will be necessary requirement
- (n) Clubbing of parameters for rationalising product offer process by Sellers.

Reverse Osmosis Equipment (Water purifier)

- (a) Type of product parameter to be added (golden parameter also) with options : Electrically Operated Reverse Osmosis based water purifier and Gravity based water purifier
- (b). Product name to be revised to generic name : **Potable water purifier Equipment**
- (c) Request made for ultra wear purifier systems (required for laboratory application) to be considered in future after obtaining inputs from stakeholders etc. for introducing same as separate product category.
- (d) Installation Type parameter to have additional option of “Ground based Installation”.
- (e) Option to choose N/A option be given against Reverse Osmosis related parameters for those offering gravity based water purifier.
- (f) Options to be added for Arsenic and Fluoride removal capability, in addition of the presently available option of Fe removal.

Patient Care Mattress:

- (a) In this item, a provision (a golden parameter also) should be provided as with or without extender which is an attachment of the mattress.
- (b) Three types of Patient Care Mattress have been identified by the seller which are in use in the market. They are : (i) Active Mattress (using compressed air pressure), (ii) Passive Mattress (using foam) and (iii) Combined Mattress (Foam + Air Pressure). But, in this case, we will continue with only Passive Mattress. Remaining types shall be considered separately after obtaining inputs from stakeholders.
- (c) A parameter should be added, i.e. Water proof of the mattress. As per specification IPX – 54 (which is to be provided by the seller as per discussion) and the water proof column has to be merged with the fire retardant and anti-microbial properties.



- (d) Mattress length, thickness and density to be standardised.
- (e) Material of mattress such as poly urethane should be added along with foam.
- (f) No. Of sections to have option 1,2,3 and 4. Parameters for seeking their lengths to be discontinued. Also the parameters related to Straps to be discontinued, since Straps itself not to continue as a parameter.
- (g) Also weight of mattress to be deleted, after specifying the mattress density parameter.

Partition Screen

- (a) No of partitions to be standardised with options 1,2 or 3
- (b) Over all length and height parameter will have standardised option(s) to select from
Moveable /Fixed frame tube dia and thickness shall be ≥ 25 mm and 1.2 mm or more
- (c) Moveable/Fixed frame tube materia shall have options of MS Tube and MS CRC tube
- (d) Leg frame dia and thickness shall be ≥ 30 mm and ≥ 1.2 mm respectively
- (e) Finish of frame shall be EPC powder quoted
- (f) Design of hinge parameter to be discontinued
- (g) Color of frame shall have options white, Ivory, Grey and Blue
- (h) Screen fabric shall have options of Cotton, Polyesterene blended cotton
- (i) Colour of frantic shall have options of Blue and green
- (j) Caster wheel size shall be 50 mm or more.

No suggestions were received regarding **Minor Procedure Light for Hospital** and **Walking Stick** during the consultative meeting.

Regards

GeM-Admin