

## Consultative Meeting Notification 45/2017

**Buyer-seller consultation on MUCUS EXTRACTORS dated  
06.12.2017**

Presented by :

Date :



## Consultative Meeting Notification 45/2017

### **Buyer-seller consultation on MUCUS EXTRACTORS dated 06.12.2017**

Consultative Meeting held on 06<sup>th</sup> December 2017 for with sellers and buyers. Following participated in the meeting:

#### **Sellers Participants**

1. M/s HOSPIMAX
2. M/s AMBU
3. M/s Track Manufacturing
4. M/s La-med healthcare (P) Ltd
5. M/s Hansraj Nayyar Medical
6. M/s Syana Health care
7. M/s Medexcel impex pvt. Ltd.

#### **Buyers Participants**

1. AIIMS
2. SAFDARJUNG HOSPITAL
3. RML HOSPITAL
4. LHMC & SSKH
5. SDDMSC
6. LBS HOSPITALS

1. Following **Golden parameters** were decided based on the suggestions and discussions –

- a) Material of Container
- b) Capacity of Container
- c) Length of Suction tube
- d) External Diameter of suction tube
- e) Filter with the container
- f) Built-in hydrophobic membrane in cap
- g) Bacterial barrier in filter in cap
- h) Method of sterilization

2. Following changes were decided in the existing TP based on the suggestions and discussions-

- a) Material of container - Non-toxic, Medical Grade
- b) Material of Tube - Non-toxic, Medical Grade – Poly Vinyl Chloride (PVC)
- c) Method of Sterilization- ETO, Gamma and Others is to be added.
- d) Tissue implantation Testing certificate– Mandatory



- e) Certification – Must Declare ISO/European CE/US-FDA
  - f) Certificate to be provided to the Buyer on demand
-