

Consultative Meeting Notification 9/2017

Buyer-seller consultation on Nebulizer dated 22-11- 2017

Presented by :

Date :





Consultative Meeting held on 22 rd November 2017 for **Nebulizer** with sellers and buyers.

Following participated in the meeting:

Sellers participants:

- 1) M/s 3m india LTD
- 2) M/s Brij Textiles
- 3) M/s Medicare product inc
- 4) M/s Durga surgical ltd
- 5) M/s kshemin consultants
- 6) M/s swarn singh khurana & sans
- 7) M/s v.k. textiles .
- 8) M/s areva international.
- 9) M/s paseca Healthcare LLP
- 10) M/s Syana India.
- 11) M/s BSN Medical
- 12) M/s Medical Devices Pvt Ltd
- 13) M/s Seven seas
- 14) M/s I Life medical devices pvt ltd
- 15) M/s Kanam Latex Ind. pvt Ltd.

Buyers participants:

- 1) Lok Nayak Hospital
- 2) SDDMASC
- 3) ESI Hospital
- 4) ACHARYA SHREE BHIKSHU GOVT HOSPITAL
- 5) LHMC AND ASSOC. HOSPITAL
- 6) M/O Health & Family Welfare
- 7) Safdarjang Hospital
- 8) AIIMS, New delhi



1. Following **Golden parameters** were decided based on the suggestions and discussions –

- i. Type- Nebulizer with/without Battery back up
- ii. Air Flow in Litre Per Minute
- iii. Nebulising Rate
- iv. Granule Size of medication
- v. Size of uniform Particle (MMAD)
- vi. Medication capacity
- vii. Hours of Battery Back up
- viii. Weight
- ix. Maximum air Flow
- x. Noise level

2. Following changes were decided based on the suggestions and discussions-

- ☐ Type should include only Nebulizer as against Nabulizer with Battery backup(In built inverter) and with/without battery back may be added as separate parameter
- ☐ “Ultrasonic” may also be added under technology parameter
- ☐ There should be separate parameter for declaration of product certification as BIS / European-CE / US-FDA / CE and manufacturing certification as ISO:13485 2003 , ISO 14001: 2004 ,OHSAS 18001 accreditations
- ☐
- ☐ **“Availability of Test Reports from Central Govt./NABL/IL AC accredited Lab “ to be replaced with "Submission of Test Report in Form 39 or from Central Govt./NABL/ILAC accredited Lab at the time of supply"**
- ☐ The details of test report like Test Report No., Test Report Date, Name of the Lab, Address of the Lab are to be deleted from the TP



GeM
Government
e Marketplace