


## Consultative Meeting Notification 143/2017

# Buyer-Seller Consultation on Spinal Needle Dated 19-12-2017



Presented by:

Date:



#### Consultative Meeting Notification 143/2017

### Buyer-seller consultation on SPINAL NEEDLE dated 19-12-2017

Consultative Meeting held on  $19^{th}$  December for SPINAL NEEDLE with sellers and buyers. Following participated in the meeting:

#### **Buyer's participants:**

- 1. AIIMS, New Delhi
- 2. RML Hospital
- 3. LHMC & SSKH
- 4. Defence, Assam Rifles
- 5. GNCTD Bhagwan Mahavir Hospital
- 6. Deepchandra Hospital

### Seller's participants:

- 1. Medax/Hospitek marketing
- 2. Kshemin Consultant
- 3. Salveo Medical Systems
- 4. Axio Biosolution
- 5. Axio Biosolution pvt. Ltd. Ahmedabad,OEM
- 6. Spanco Industries
- 7. Vaishnvi Med. Eq. System
- 8. BVM Medical System
- 9. Swastik Traders
- 1. Following **Golden parameters** were decided based on the suggestions and discussions:-
  - Outside diameter of cannula
  - Length of cannula
  - Introducer
  - Hub
  - · Length of cannula
  - introducer
- 2. Following changes were decided based on the suggestions and discussions-
  - Designation (mm) to be deleted
  - Length of cannula—120, 150 to be added



- Size of needle in gauge—18, 20, 22, 23, 24, 25, 26, 27, 28, 29, 30 to be added
- New parameter of introducer may be added
- Point of needle—quincke bevel/pencil point may be replaced for short bevel 30/45 to axis
- Smooth movement and well fitting of stilette must be made mandatory
- Method of sterilization may be ETO/Gamma
- Manufacturers Drug Licence number and date to be included
- Shelf life of 5 years to be included
- Availability of test report from Central Govt./ NABL/ILAC may be replaced with submission of test report on Form No.39 or from Central Govt./ NABL/ILAC along with supply to the buyer
- Details of test report may be deleted.

Regards

GeM-Admin