

## Consultative Meeting Notification 26/2017

Buyer-seller consultation Toomey/irrigation syringe dated 24-11- 2017

Presented by :

Date :





Minutes of the Consultative meeting held on 24.11.2017 at 11.30 PM under Chairmanship of Shri A V Muralidharan Dy CEO (GeM) at the Office of Government e Market (SPV New Delhi regarding Medical items;

**The following attended Meeting from seller side:**

1. Ms Life Long Meditech pvt. Ltd.
2. M/s Hindustan Syringe Ltd
- 3 . M/s B. Braun Medical (I) Pvt Ltd
4. M/s B.Brawn Medical (I) Pvt Ltd
5. M/s Mediplus India Ltd
6. M/s Polymedicure Medical Devices
7. M/s Sterimed Medical Devices Pvt Ltd
8. M/s ISCON Surgicals Ltd
9. M/s Synaindia associates
10. M/s Alliance Medi Equipment Ltd

**Buyer side:**

None Present.

**Suggestions received in the meeting from Participants and decided to be included in the TP**

- ☐ Intended use as stomach and bladder irrigation
- ☐ Indian Standard IS 3237(Part6):1986 and option of any other Standards
- ☐ Capacity of syringe as 20 ml, 60ml and 80 ml or others to be declared
- ☐ Length of syringe to be declared
- ☐ Declaration of Piston syringe barrel graduations
- ☐ Confirmation about piston syringe barrel fitting all standard cytosopes and reserosopes
- ☐ Drug License Provisions
- ☐ Shelf life to be declared by seller with confirmation of having at least 2/3 of shelf life



available at time of supplies.

- ☐ Good Manufacturing Practices (GMP) under revised Schedule -M or WHO-Good Manufacturing Practices (GMP) or any other certifications such as CE/FDA/CSA/PQS in respect of the original manufacturer
- ☐ Non conviction certificate.
- ☐ Requirement of copy of in house test report with each lot of supplies.

**Golden Parameters Proposed by Participants**

- ☐ Plunger Tip Design
- ☐ Nominal Capacity/Graduated capacity of Syringe (ml)

**Deletion from existing TP**

No deletions to be made

Regards

GeM-Admin



**GeM**  
Government  
**e Marketplace**