

Consultative Meeting Notification 48/2017

Buyer-seller consultation on Vented Endotracheal

Bougies dated 06.12.2017

Presented by :

Date :





Consultative Meeting Notification 48/2017

Buyer-seller consultation on Vented Endotracheal Bougies dated 06.12.2017

Consultative Meeting held on 06th December 2017 with sellers and buyers. Following participated in the meeting:

Sellers Participants

1. M/s HOSPIMAX
2. M/s AMBU
3. M/s Track Manufacturing
4. M/s La-med healthcare (P) Ltd
5. M/s Hansraj Nayyar Medical
6. M/s Syana Health care
7. M/s Medexcel impex pvt. Ltd.

Buyers Participants

1. AIIMS
2. SAFDARJUNG HOSPITAL
3. RML HOSPITAL
4. LHMC & SSKH
5. SDDMSC
6. LBS HOSPITALS

1. Following **Golden parameters** were decided based on the suggestions and discussions –
 - a) Category- Adult or Pediatric
 - b) Material
 - c) Inner Diameter (I.D)
 - d) Method of sterilization
2. Following changes were decided in the existing TP based on the suggestions and discussions-
 - a) Category – Adult or Pediatric to be added.
 - b) Length -700mm – Other sizes are to be added on receipt of information from the buyers as stated in the meeting on 06.12.2017
 - c) Material – Non-toxic Medical grade – Must be declared
 - d) Method of Sterilization- (1) ETO, (2) Gamma and (3) Others to be added.
 - e) Tissue implantation Testing certificate – Mandatory
 - f) Certification – Must Declare ISO/European CE/US-FDA
 - g) Manufacturing Drug License- Yes
 - h) Authority of Drug License- Must Declare
 - i) Drug license No. - Must Declare
 - j) Drug license Date- Must Declare



- k) The product should have shelf life of at least 2/3 rd of the total shelf life at the time of supply
 - l) All certification to be provided to the buyer on demand
-